



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY

G T Road By Pass, Jalandhar-144011, Punjab (India)
EPABX-0181-2690301 & 453 website: www.niti.ac.in email: registrar@niti.ac.in

Ref. e-Tender Notice No. NITJ/PUR/CE/176/20/ e-tender no. 34/2022

Please find enclosed herewith the following:		Page No.
1.	e-Tender Notice No. NITJ/PUR/CE/176/20/ e-tender no. 34/2022 & Schedule for Opening of the e-Tender	2
2.	Annexure-I (Detail of Equipment, Tender Fee & EMD)	3
3.	Important Note	4
4.	Instruction to Tenderers	5-7
5.	Tender Evaluation	8-10
6.	Terms & Conditions	11-12
7.	Acceptance of terms and conditions	13
8.	Questionnaires A & B	14-15
9.	Performa for performance statement	16
10.	Annexure 'A' (Specifications of the Equipment)	17-23
11.	Annexure 'B' & 'C' (Format For Performance Bond/ Guarantee & Format For Performance Bond (Bank Guarantee)	24-25
12.	Annexure – 'D' (Format for Manufacturer's Authorization Form)	26
13.	Annexure- 'E' (Declaration Regarding Blacklisting/ Debarring for taking part in Tender)	27
14.	Annexure –'F' (Certificate of Warranty)	28
15.	Annexure-'G' (Self Declaration for local content)	29

Assistant Registrar

Dr. B. R Ambedkar NIT Jalandhar

Email: arpurchase@niti.ac.in



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY

G T Road By Pass, Jalandhar-144011, Punjab (India)

EPABX-0181-2690301 & 453 website: www.nitj.ac.in email: registrar@nitj.ac.in

THIS DOCUMENT IS FOR REFERENCE ONLY. ONLY E-TENDERS WILL BE ACCEPTED

e-Tender Notice No. NITJ/PUR/CE/176/20/ e-tender no. 34/2022

National Institute of Technology, Jalandhar invites e-tender for Purchase of **Automatic Compression Testing Machine with Flexural Test Attachment – Under HEFA** required for **Structures Testing Lab of Civil Engineering Department** as per detail available at Annexure-I along with Tender Fee & EMD as per details given below:

I	Downloading & Submission of Online e-tender/bids	Start Date: 25.08.2022 at 03:00 PM
II	Last date of submission of online bids	End Date: 15 .09.2022 upto 03:00 PM
III	Physical submission of Tender Fee and EMD	End Date: 16.09.2022 upto 03:00 PM
IV	Opening of Technical e-Bid (online)	16.09.2022 at 03:00 PM

Detailed Terms and Conditions are available in e-tender document. The bid document can be downloaded from the [CPP Portal](#).

Complete tender document is available for reference purposes on Institute website www.nitj.ac.in and [CPP Portal](#). Only e-tenders will be accepted.

NOTE: This is a domestic Tender according to the DPIIT Order dated 16.09.2020 and subsequent amendments to the order for Public Procurement Preference & PROVISION FOR LOCAL SUPPLIERS TOWARDS PREFERENCE TO MAKE IN INDIA. The bidder required to provide a declaration from the Original Equipment Manufacturer (OEM) on their letter head the percentage of Local Content for the quoted instrument / item and submit with the Technical bid as per Annexure-G. Only class-I and Class –II Local suppliers will be eligible to participate in tender. Non-Local supplier are not eligible to participate in tender. Bidder should also give details of the location(s) at which the local value addition is made in the Annexure, failing which the bid be rejected.

Registrar



Annexure-I

Ref. No. e-Tender Notice No. NITJ/PUR/CE/176/20/ e-tender no. 34/2022

Detail of Equipment, Tender Fee & EMD are as under:

Sr. no.	Item/Equipment	Qty.	Tender Fee	EMD
1.	Automatic Compression Testing Machine with Flexural Test Attachment	01 no.	Rs. 500/-	Rs. 1,00,000/-

Note: The quantity of required equipment/item may vary as per requirement.

***Exemption of Tender fee & EMD will only be given to MSME/NSIC registered bidders.**



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY

G T Road By Pass, Jalandhar-144011, Punjab (India)

EPABX-0181-2690301 & 453 website: www.nitj.ac.in email: registrar@nitj.ac.in

Important Note

1. All corrigenda, addenda, amendments and clarifications regarding this tender document will be uploaded on the website www.nitj.ac.in and CPP Portal and not in the newspaper; Bidders shall keep themselves updated with all such developments.
 2. In case, the last date of receipt/opening of bids falls on holiday, the bids shall be receipt/opened on the next working day at same time.
 3. In case, the last date of submission of EMD & Tender fee falls on holiday, the EMD & Tender fee shall be submitted on the next working day at same time.
 4. Tenderer who have downloaded the tender document form from the institute website, shall submit a declaration along with tender document that I/We have downloaded the Tender Form from the institute website www.nitj.ac.in and I/we have not tempered /modified the tender form in any manner. In case, if the same is found to be tempered/modified in any manner, I/we understand that my/our tender will be summarily rejected and I/we are liable to be banned from doing business with institute.
 5. **Tender fee of Rs.500/- (Non- refundable) in the form of DD in favour of Director, D R B R Ambedkar NIT, Jalandhar.**
 6. **EMD (refundable) in the form of DD in favour of Security- A/c. DR B R Ambedkar NIT, Jalandhar.**
 7. **Both EMD and Tender fee are be submitted as per dates mentioned in schedule, failing which e-bids will not considered.**
- ❖ **All the bidders are required to submit the Tender Fee and EMD as per requirement of tender document failing which bids received straightway rejected and bid will be treated invalid.**
 - ❖ **Note: If the bidder inadvertently or otherwise upload the quoted rates in the technical bid, the bid will be straightway rejected and treated invalid.**
 - ❖ **If the bidder is exempted for payment of Tender Fee and EMD as NSIC/MSME registered bidders, then bidder is required to submit NSIC/MSME exemption certificate for same. The Certificate must be valid as on last date of submission of bid.**

Tenderer must submit a scanned copy (duly signed and stamped) regarding terms & conditions as per our tender documents along-with make/model, specifications, bill of quantity as per required equipment in the technical bid for examine the bid as per our institute tender documents. It is noted that no rate should be depicted in the letter head.



Instructions to Tenderer

1. No tender will be accepted in physical form. The bidders shall have to submit their bids online in Electronic Format under Digital Signatures. For participation in the e-tendering process, the bidders need to register themselves on CPP Portal.
2. Bids are to be submitted online and opened online as per time given failing which no tender will be considered.
3. Bids will be opened online as per time given schedule.
4. **Before submission of online bids, bidders must ensure that scanned copies of all the necessary/relevant documents have been uploaded with the bid which should be duly signed and stamped. The duly signed and stamped copies of Terms & Conditions of the tender, reply of the Questionnaire of Plant & Machinery and other documents of the Tender & Annexures must be uploaded, failing which their bids may be rejected.**
5. NIT JALANDHAR, will not be responsible for any delay in online submission of bids due to any reason whatsoever.
6. **Bidders should also upload the scanned copies of Tender fees/EMD/Exemption Certificate as specified in the tender documents along with online technical documents. EMD in the form of a Demand Draft in favour of the Security – A/c, Dr B R Ambedkar NIT, payable at Jalandhar (refundable separate) and Tender Fee in the form of a Demand Draft in favour of the Director, Dr B R Ambedkar NIT, payable at Jalandhar (Non- refundable separate) should also be submitted in physical form to the following address as per scheduled time given for physical submission of EMD and Tender fee. The Envelope should be super-scribed as EMD and Tender Fee for Tender for Supply of Automatic Compression testing Machine with Flexural Test Attachment of Civil Engineering Department of the Institute and sent to following address:-**

**Kind Attention- Assistant Registrar (Purchase Section)
Director,
Dr B R Ambedkar National Institute of Technology,
G T Road Amritsar By Pass, Jalandhar-144027, Punjab (India).**
7. The details of EMD specified in the tender document should be same as submitted online (scanned copies). Otherwise tender will be rejected summarily.
8. The conditional bids shall not be considered and will be out rightly rejected.

Read and Accepted

(Signature & Stamp of Tenderer)



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY

G T Road By Pass, Jalandhar-144011, Punjab (India)

EPABX-0181-2690301 & 453 website: www.nitj.ac.in email: registrar@nitj.ac.in

9. The Financial Bid through e-tendering process shall be opened of only those bidders, who will qualify in the technical bid and approved by the Purchase Committee/Technical Experts. The date, time & place of opening of the financial bid(s) will be intimated in due course of time.
10. At any time prior to the deadline for submission of bid, the institute may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective tenderer(s), modify the tender document by issuance of an amendment.
11. The amendment will be uploaded on Institute website and CPP Portal only. In order to provide reasonable time to prospective tenderer(s), for preparing their bid as per amendment, the institute may, at its discretion extend the deadline for the submission of tender.
- 12. The supplier must upload the original manuals / catalogue and Make/Model of the Equipment /Item. Otherwise bid is liable to be rejected.**
13. The Institute is not liable to pay any interest on EMD. Earnest money deposit shall be forfeited, if the tenderer, withdraws its bid during the period of tender validity. The Earnest money deposit of the tenderer, whose tender has been accepted, will be returned on the submission of **performance security @ 3% of the total value of the offer. The performance security will be kept till the warranty period + 02 months more of the Equipment /Item. The warranty period will start from the date of satisfactory installation of the Equipment /Item duly given by the concerned department.** Earnest money deposit of the successful tenderer shall be forfeited, if it refuses or neglects to execute the contract or fails to furnish the required performance security within the time frame as specified by the institute. The EMD(s) of other Bidder(s) whose offer are found according to required specifications/lowest will be released after finalization of Technical Bids/ Lowest Bid/Purchase.
14. The Format of Performance Bank Guarantee bond or Performance Bank Guarantee issued by the bank as per the format given in **Annexure "B" & "C"**.
15. Delivery time is the essence of the contract and must be met with.
16. Nearest specifications/better specifications can be considered. In case of deviation, complete justification should be furnished with proper documents.
17. The Director may accept a tender in part or whole of the quantity offered, reject any tender without assigning any reasons and may not accept the lowest bidder. Further in case of any doubt/dispute, the decision of the Director of the Institute shall be final.
18. The offer shall be kept valid for minimum 120 days.

Read and Accepted.

(Signature & Stamp of Tenderer)



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY

G T Road By Pass, Jalandhar-144011, Punjab (India)

EPABX-0181-2690301 & 453 website: www.niti.ac.in email: registrar@niti.ac.in

19. a) Either the agent/ dealer on behalf of the Principal/OEM or Principal/OEM itself can bid but both cannot bid simultaneously for the same item/product in the same tender.
- b) If an agent/ dealer submits bid on behalf of the Principal/OEM, the same agent /dealer shall not submit a bid on behalf of another Principal/OEM in the same tender for the same item/product.
- c) Only one agent/dealer of single Principle/OEM can submit the bid. If more than one agent/dealer of same OEM/Principal quote in the tender, then bids of all the bidders of single OEM/Principle are liable to be rejected.
- d) All offers other than those from the Principal/OEM should be supported by an authority letter from the manufacturer authorizing the dealer /supplier to tender on their behalf as per **Annexure-D**. In case of manufacturer, a certificate or a copy thereof to the effect that the bidder is a manufacturer of the Equipment /Item must be accompanied with the technical bid.
20. The supplier will be responsible till the entire stores contracted for, arrive in good condition at destination.
21. The tenderer should not have been debarred and/ or blacklisted by any Central Government/ or any State Government Department(s) . This must be supported by an affidavit as per format given in **Annexure-“E”**.
22. If any information furnished by the bidder is, at any stage found to be incorrect/false/fabricated, the Institute shall have the absolute right to forfeit the EMD, warranty/performance guarantees or/and security deposits, in addition to cancellation of contract, and in accordance with law, such other actions may be taken like black-listing of the bidder etc.

Read and Accepted.

(Signature & Stamp of Tenderer)



TENDER EVALUATION

Institute will evaluate all the proposals to determine whether these are complete in all respects as specified in the tender document. Evaluation of the proposals shall be done in two stages as:

(a) Stage – I (Technical Evaluation):

- (i) Institute shall evaluate the technical bid(s) to determine the following like the bid qualifies the essential eligibility criteria or not, the tenderer has submitted the EMD & Tender fee or not, any computational errors have been made or not, all the documents have been properly filled or otherwise, all the documents have been submitted/ uploaded with technical bid or not, the specifications, Make/Model, Catalogue of quoted Equipment /Item are as per requirement tender specifications or not, Authorization of Dealer / Distributor/ Exclusive Agent certificate from manufacturer is in order or not, Sales & service policy of equipment / item during warranty period and after warranty period will also be seen, location of their authorized service center will also be seen for evaluation etc.

After evaluation of technical bid(s), a list of the qualifying tenderer (s)/ bidder (s) shall be made. Short-listed tenderer(s) will be informed of the date, time and place of opening of financial bid(s) and they may attend or depute their authorized representative/s to attend the schedule of opening of financial bid(s) on the scheduled date and time, if they wish to do so. The representative(s) should have a letter of authority to attend the price bid(s) opening event.

Read and Accepted

(Signature & Stamp of Tenderer)



PART – II (e-FINANCIAL BID):

- a) Bidders should offer the rates as per the format of BOQ as available on CPP Portal. **Detailed bill of material/quantity is also to be provided along with the price breakup of each item as per requirement of the tendered specification of the equipment in the online price bid at [CPP Portal](#).**

❖ **Note: The quoted amount as filled in the Annexure(s) of online financial bid and detailed bill of material/quantity provided with price break up of each item in the online financial bid should be tallied and both must be same, otherwise bid will be treated invalid.**

Read and Accepted.

(Signature & Stamp of Tenderer)



ARBITRATION:

In case of any dispute or difference arising out in connection with the tender conditions/job order/Contract, the Institute and the Seller/Service Provider will address the dispute/difference for a mutual resolution and failing which, the matter shall be referred for arbitration to a sole Arbitrator to be appointed by the Institute.

The Arbitration shall be held in accordance with the provisions of the Arbitration and Conciliation Act, 1996 and the venue of arbitration shall be at Jalandhar only. The decision of the Arbitrator shall be final and binding on both the parties.

JURISDICTION:

The courts at Jalandhar alone will have the jurisdiction to trial any matter, dispute or reference between parties arising out of this tender / contract. It is specifically agreed that no court outside and other than Jalandhar Court shall have jurisdiction in the matter.

Read and Accepted.

(Signature & Stamp of Tenderer)



(TERMS AND CONDITIONS (FOR THE SUPPLY OF GOODS, EQUIPMENT /ITEM)

1. Rate should be quoted F.O.R NIT Jalandhar and In INR only.
2. Where the Equipment / Machinery/ Instrument are composed of several sub units/components, the rate should be quoted for each subunit/component separately. NIT Jalandhar reserves the right to increase or decrease the number of sub units/components and number of Equipment / Machinery/ Instrument according to its requirements.
3. The institute being government educational institute is having Excise and Custom duty exemption in terms of government notification No. 51/96-customs Dt. 23-7-1996 and No. 10/97-central excise Dt. 1-3- 1997 as amended from time to time, therefore taxes be quoted accordingly and this must be depicted in Price Bid clearly.
4. GST or any other chargeable duty where applicable must be specifically mentioned, failing which no tax or duty will be allowed at subsequent stage.
5. All items shall be indicated both in words as well as in figures. If there is difference between amount quoted in words and figures, amount quoted in words shall prevail.
6. **Payment:**
 - (a) 100% payment will be made against physical delivery, inspection, installation, training of the Equipment/Machinery/Instrument etc. in the institute, receipt of satisfactory working report of the Equipment / Machinery/Instrument etc. and receipt of Performance Bank Guarantee @ 3%.
7. **Warranty: Rates should be quoted with comprehensive warranty.** Besides this, policy regarding after sale service on the expiry of warranty period of Equipment / Machinery/Instrument etc. may be explained. The bidders should attach duly signed and stamped certificate of **warranty** as per **Annexure-F** with the technical bid.
8. **Training: In house training (where applicable) after the installation and commissioning of Equipment / Machinery/Instrument etc. shall be provided by the supplier.**
9. **Delivery:** Delivery date will be mentioned in the supply order. The time and date of delivery or dispatch stipulated in a supply order shall be deemed to be the essence of the supply order and if the supplier fails to deliver or dispatch any consignment within the period prescribed for such delivery, the delayed consignment will be accepted subject to penalty as laid down in the supply order, which will be recovered from the pending payments.

Read and Accepted.

(Signature & Stamp of Tenderer)



1. No recovery of penalty will be made, if the delayed supplies are acceptable by extending the delivery period by the Director with our any LD charges.
2. Director will allow extension on the request of the supplier by recording in writing that in exceptional circumstances the supply was beyond the control of the supplier and there was no loss to the institute.
3. Penalty on account of delay, Director NIT, Jalandhar reserves the right to impose 0.5% (Half) per cent penalty per week on account of delay in supply, if delivery received after expiry of the original delivery period. The total penalty will not exceed 10% of the value of the delayed goods.

10. Installation: Supplier has to install the Equipment / Machinery/ Instrument within two to three weeks from the receipt of the Equipment / Machinery/Instrument etc. in NIT Jalandhar.

11. Spares and Accessories, wherever required should be quoted separately and clearly, even if these are not asked for.

12. Site Preparation: The supplier shall inform NIT Jalandhar about the site preparation, if any, needed for the installation, immediately after receipt of the supply order. Supplier must provide complete details regarding space and all infrastructural requirements needed for the Equipment / Machinery/Instrument etc which NIT Jalandhar should arrange before the arrival of Equipment / Machinery/Instrument etc to ensure its early installation and smooth operation thereafter. The supplier may offer his advice and render assistance to NIT Jalandhar in the preparation of the site and other pre installation requirements.

13. The total scope of work includes the supply, installation, satisfactory commissioning and testing of the Equipment / Machinery/Instrument etc by the supplier, training at NIT Jalandhar, method development and validation for parameters as mentioned in specifications at **Annexure-A**. The supplier will complete installation & Commissioning of Machine within two to three weeks from the date of receipt of Equipment / Machinery/Instrument etc. of NITJ.

14. Details about the service center for the quoted Equipment / Machinery/Instrument etc. in India may be mentioned.

Read and Accepted.

(Signature & Stamp of Tenderer)



Acceptance

We _____ read and accept the instructions to the tenderer, terms & conditions and all other documents as mentioned in the tender and shall comply with them strictly.

Name of Bidder _____

Signature

Address _____

Seal of firm:

Date:



Questionnaires A & B

QUESTIONNAIRE FOR PLANT & MACHINERY

Note: Please submit the reply in detail and also enclose the necessary documents with proof where required as per Questionnaire. These documents must be submitted in the technical bid document.

Please state that you have submitted your quotations as per procedure mentioned below:-

➤ Bids consisting of technical details bringing out clearly in a separate sheet, the deviation in the specifications if any from that of tender enquiry specifications along with commercial terms and tender form.

1. Please attach list duly signed by you, for such spare parts and tools which are absolutely essential for proper maintenance and operation of machine for a period of two years giving full particulars of spare and tool with a price of each spare parts and tools separately.
2. Please confirm that you have adequate servicing and spare parts facilities in India in respect of Equipment /Item tendered by you or that you should arrange to provide such facilities simultaneously.
3. Please indicate that you guaranteed that before going out of production of spare parts, you will give adequate advance notice to the purchaser so that the institute may order his requirements of spares in one lot, if he so desire.
4. Please indicate that you guaranteed that if you go out of production of spare parts, then you will make available prints, drawings of the spare parts and specifications of the material at no cost if and when required in connection with Equipment /Item to enable the purchaser to fabricate or procure spare parts from other sources.
5. Please confirm that you undertake to enter into a rate contract with the purchaser to supply spare parts on an agreed basis for an agreed period.
6. Status:-
 - a) Indicate whether you are ISU or SSI
 - b) Are you registered with Government e-Marketing (GeM) for the item quoted? If so indicate whether there is any monetary limit or registration.
 - c) If you are a small scale unit registered with NSIC under single point registration scheme whether there is monetary limit.
7. a) If you are registered either with NSIC or with Government e-Marketing (GeM), please State whether you are registered with Directorate of industries of the state government concerned.

b) If so, confirm whether you have attached a copy of the certificate, issued by Director of industries.
8. Please indicate:-
Name & Full Address of your Banker
9. Please indicate whether you agree to submit advance samples if called upon to do so within the specified period of 21 days.



10. Business name and constitution of firm:-

- i) The Indian company Act. 1956
- ii) Indian partnership Act. 1932
- iii) Any act, if not, who are the owner/partners (please give full name and address)

11. Whether the tendering firm is / are:-

- i) Manufacturers authorized agents
- ii) Holders stock of the stores tendered for
- iii) **NBI manufacturer's agents please enclose with tender the copy of manufacturer's authorization.**

12. Please state whether the inspection clause is acceptable to you

13. Here state specifically:-

- i) Whether the price quoted by you is to the best of your knowledge and belief is not more than the price usually charged by you on stores of the same natures, class of description to any private purchase either foreign or as well as government purchaser. If not state the reason thereof if any also indicates the margin of Difference.
- ii) In respect of indigenous items for which there is a controlled price fixed by law, the price quoted shall not be higher than that the controlled price and rates available on DGS&D/ Government e-Marketing (GeM) contract. The reason thereof should be stated.

14. State whether business dealing with you has been banned by Ministry /Department of supply or any other Govt. Department.

15. Please confirm that you have read all the instruction carefully and have complied with accordingly.

(Signature of Tenderer)

(_____)

1. Full Name & Address of the person signing

(_____)

2. Whether signing as proprietor /partners



PERFORMA FOR PERFORMANCE STATEMENT

(For the Period of last 3 years)

e-Tender Notice No. NITJ/PUR/CE/176/20/ e-tender no. 34/2022						
Sr. No.	Order (full placed address of purchaser)	Order No. & dated	Description And Qty. of Stores Ordered	Value of Order	Date of Completion of delivery	Has the Equipment/Item been satisfactory commissioned



Subject: Purchase of 'Automatic Compression Testing Machine with Flexural Test Attachment' approved under HEFA.

The equipment shall include all components necessary for complete system operation including set of high stiffness test frames, extensometer, auto-control unit, data acquisition & relevant software along with the compatible personal computer.

The equipment should comprise of the following:

S No	Item	No.	Description
1	Computerized Control Unit	1	Servo-hydraulic Unit and Controller
2	Compression Test Frame	2	5000 kN, 500 kN
3	Flexural Test Frame	1	350 kN

Important Note: Please quote separately for each major component/part/accessory thus enabling us to select as per requirement. However, the offers shall be evaluated on the basis of total cost.

1. Technical specifications for the Computerized Control Unit

Hydraulics:

- Should be capable of implementing servo-hydraulic control of load, displacement and strain
- Should be capable of producing a continuous and non-pulsating oil flow of suitable pressure
- Should have the provision of connecting to a minimum of four test-frames and run any one at a given time
- Should be capable to connect 15 kN to 5000 kN compression frame and up to 500 kN tensile testing frame.
- Should be capable of running on standard AC single or three phase supply -operating under the ambient temperature up to 40°C without any external cooling requirements
- The noise rating of hydraulic power pack should be lesser than 75 db
- The hydraulic power pack should be vibration free and compact
- Minimum working pressure should be 650 bar



Controller:

- Should be based on a PID algorithm with an update rate of 500 Hz or better
- Should take user specified values of test control parameters - load, displacement and strain rates - through software GUI as set points
- Should be capable of driving the hydraulics to generate maximum and minimum loading rates of 15 kN/s and 0.01 kN/s respectively or better
- Should be capable of driving the hydraulics to generate maximum and minimum displacement rates of 0.09 mm/s and 0.018 mm/s respectively or better
- Should be capable of providing a load resolution of 0.1 kN or better at an accuracy of $\pm 1\%$ or better
- Should be capable of providing a displacement resolution of 0.001 mm or better at an accuracy of $\pm 1\%$ or better
- Should have a minimum of 12 input channels for data acquisition through load, displacement and strain transducers. Each channel should be calibrated for flexible inputs
- The data acquisition hardware should incorporate essential circuitry, chips and buses for signal conditioning, conversion and transmission. The sampling of analogue signals should be at a rate of 100 Hz or faster and corresponding digital signals should be generated at a resolution of 16 bits or better
- Should be capable of automatically turning the pump on and off, controlling the set pace rate and switching off the machine under predetermined conditions of load, displacement or strain
- Should be capable of automatically releasing the pressure at the end of every run and re-initialize the machine at the beginning of every test

PC interface:

- Should be from the original equipment manufacturer (like DELL, HP, Lenovo or equivalent) with at least 10 years of history in manufacturing personal computers. It should not be an assembled computer with components from various manufacturers. Along with appropriate capacity UPS.



Associated software:

- Compatible with Windows operating system
- Should have a GUI with input fields for: loading, strain and displacement rates; specimen shape and size
- Capable of active frame selection
- Capable of real-time data acquisition, processing and display of load-time and stress-strain curves
- Capable of automatic verification of sample position and sensors functionality, as per standards requirements
- Capable of archiving data in text or spreadsheet format
- Capable of printing test reports
- Should include data manager software package for compression, indirect tensile, 3 points and 4 points flexural tests on different types of specimens
- Capability to determine the Elastic modulus and Poisson ratio
- Capable of automatic calculation of test results conform to EN 14651, 14488-3, 14488-5, UNI 11039-2, ASTM C1550, C1609, C947, UNE 83515
- Capability of changing the real-time test parameters such as target load/displacement, control variables and test speed

2. Technical specifications for Compression Test Frames

Capacity: 5000 kN

- The frame should have a column/plate type structure with safety guard doors
- The frame should have a single acting hydraulic cylinder at its centre
- The frame should have a longitudinal stiffness of 4 MN/mm or better*
- The loading platens should have a Rockwell hardness of 53-60 HRC or better
- The top platen should have a spherical ball seat arrangement and the bottom platen should have suitable marking to facilitate the centering of cubic and cylindrical specimen
- Platen size: 300 mm or bigger
- Minimum vertical clearance between platens: 600 ± 20 mm
- Minimum horizontal clearance: 425 ± 20 mm
- Suitable distance pieces should be provided to facilitate holding of specimen of up to 100 mm height



- The actuator should have a suitable transducer capable of measuring load in range of 1% to 100% of the full-scale with a resolution of 0.1 kN and an accuracy of $\pm 1\%$ or better.
- The actuator should have a suitable transducer capable of measuring displacement of up to 50 mm at 0.001 mm resolution with an accuracy of $\pm 1\%$ or better
- Maximum travel of ram should be equal to 50 mm with suitable travel limit switch
- The frame should conform to the force transfer requirements as per EN 12390-4**
- The calibration should conform to IS 516, EN 12390-4, ASTM C39 and AASHTO T22
- The precise loading frame stiffness calculation is mandatory and the same shall be provided to NIT Jalandhar for verification.
- A certificate validating the uniform force transfers should be submitted and will be considered for technical compliance.

Capacity: 500 kN

- The frame should have a plate type structure with safety guard doors
- The frame should have a single acting hydraulic cylinder at its centre
- The loading platens should have a Rockwell hardness of 53-60 HRC or better
- The top platen should have a spherical ball seat arrangement and the bottom platen should have suitable marking to facilitate the centring of cubic and cylindrical specimen
- Platen size: 150 mm or bigger
- Minimum vertical clearance between platens: 300 mm or above
- Minimum horizontal clearance: 260 mm or above
- The frame should be capable of holding cubic specimen of up to 70 mm height, if required, through suitable accessories
- Capacity: 25 KN to 500 kN; Min. resolution: 0.1 kN; Accuracy: $\pm 1\%$ or better
- Max. limit: 50 mm; Min. limit: 0.001 mm; Accuracy: $\pm 1\%$ or better



3. Technical specifications for Flexure Test Frame

Capacity: 350 kN

- The frame should have a parallel and orthogonal testing facility
 - The frame should have a single acting hydraulic cylinder at its centre
 - The effective stiffness of the frame should be 175 kN/mm or better*
 - The frame should have the capability of conducting flexural toughness tests on strain softening materials as per ASTM 1609, ACI 544, EN 14651, ASTM 1550 and IRC SP46 and regular flexural strength tests as per IS 516
 - The frame should be capable of implementing three point and four point flexural tests
 - The frame should be capable of testing beam specimens of 0.5 to 1.5 m length or longer
 - Minimum vertical clearance: 400 mm or above
 - Minimum horizontal clearance: 900 mm or above
 - The frame should have adjustable vertical clearance along with spacers
 - Distance between loading rollers: Adjustable from 100 mm to 500 mm or higher over a suitable scale
 - Distance between support rollers (Parallel test): Adjustable from 200 mm to 800 mm or higher over a suitable scale
 - Distance between support rollers (Orthogonal test): Adjustable from 800 mm to 1500 mm or higher over a suitable scale
 - Accessories for loading and supporting the specimen should suit the modes of testing conforming to various standards such as IS 516, ASTM 1609, ACI 544, EN 14651, ASTM 1550 and IRC SP46
 - The support rollers should ensure simply supported condition
 - While testing under four-point loading, the loading head should swivel, along the axis parallel to the longitudinal axis of the rollers, to ensure complete contact with the specimen surface throughout the loading period
 - During testing under three-point loading configuration, the arrangement should allow locking the loading head in position arresting the swivelling motion
 - Max. Capacity: 350 kN; Min. resolution: 0.1 kN; Accuracy: $\pm 1\%$ or better
 - Displacement transducer (LVDT type): Max. limit: 100 mm; Min. limit: 0.001 mm; Accuracy: $\pm 1\%$ or better
- * The precise loading frame stiffness calculation is mandatory and the same shall be provided to NIT Jalandhar for verification.



Accessories:

- CMOD-Displacement transducer for measurement of crack opening (Quantity – 1)
 1. The minimum sensitivity should be 2.5 mV/V
 2. With fixing jigs
 3. Conform to EN 14651
- Electronic extensometer (Quantity -3)
 1. Capable of testing standard cylindrical specimens as per IS 516
 2. Capable of testing prism specimen with max. length of 160 mm
 3. Accuracy on strain measurement: 2×10^{-6} or better
 4. Should supply along with adjusting template and elastic band
 5. Conforming to EN 12390-13, EN 13412:2006, ASTM C 469, ISO 1920-10:2010
- Cylinder compressometer/extensometer system (Quantity -1)
 1. Capable of testing standard cylindrical specimens as per IS 516
 2. Displacement transducer should be capable of measuring of up to 10 mm or better (resolution 0.001 mm and an accuracy of $\pm 1\%$ or better)
- Specimen holder for split tensile test (Quantity – 1)
 1. Cylindrical specimen with a diameter in the range of 100 to 200 mm and up to 300 mm length: Conform to EN 12390-6:2009, ASTM C 496
- Accessories for the flexural toughness of circular slab (Quantity - 1)
 1. Three point spherical seats
 2. Maximum test specimen diameter: 800 mm
 3. Conform to ASTM C 1550-12a
- Accessories for the Energy absorption test on square slab (Quantity - 1)
 1. Square supporting frame ensuring simply support condition
 2. Conform to EN 14488-5 and UNI 10834
- Accessories for the flexural testing of fiber reinforced concrete (Quantity - 1)
 1. Auxiliary testing frames capable of measuring displacement and holding beam specimens of following sizes:
 - a) 100 mm \times 100mm \times 500 mm
 - b) 150 mm \times 150mm \times 700 mm
 2. Displacement transducer (LVDT type): 2 Nos.
Max. limit: 10 mm; Min. limit: 0.001 mm; Accuracy: $\pm 1\%$ or better
 3. Electrical mean device for 3 displacement transducers
 4. Conforming to BS EN 14488-3:2006, ASTM C 1609



48-29

Elastomeric bearing sheets (Quantity - 1 set)

The suitable elastomeric bearing sheets should be provided to support the equipment, which serves as an intermediate layer between the floor and equipment.

Note: If any additional optional accessories are required

Other Terms:

- 1 Original Printed catalogue with model numbers must be submitted along with the bids. Just typing tender specification on company letterhead will not be considered as a valid document.
- 2 The bidder must have done three installations and commissioning of 5000 kN or higher capacity compression testing machine/s conforming to specifications, with traceable calibration certificate at IITs/NITs institutions. The documentary evidences viz., installation and commissioning certificate/performance certificate, copy of calibration certificate for 5000 kN compression testing machine/s and purchase order copy/s, all in English language must be submitted as a part of the technical bid.
- 3 After complete installation, at least 3 days of onsite training shall be provided to students/staff/faculty. Installation of machine and training must be provided on machine operation, data interpretation & analysis.
- 4 A minimum 2 years of full warranty shall be provided for complete system including accessories.
- 5 From the date of installation, all the spare parts should be available for at least 10 years. The vendor shall supply as and when order is placed for repair/replacement.



Annexure: “B”

FORMAT FOR PERFORMANCE BOND/GUARANTEE

(Undertaking from the supplier on a Non Judicial Stamp Paper of requisite duly attested by Notary)
In consideration for “The Registrar, National Institute of Technology Jalandhar (hereinafter called NIT Jalandhar) having agreed to release the payment of net value as per terms and conditions of a concluded Order No. _____ dated _____ (hereinafter called ‘the order’) for supply of _____ (here in after called ‘the Equipment /Item)to us Messrs _____ (hereinafter called ‘the supplier’) on submission of a Performance Bond to the satisfaction of NIT Jalandhar for the due performance of the said order.

We, Messer’s _____ hereby submit the FDR/TDR No. _____ issued by _____ (Name of Bank) for _____ pledged in favour of Registrar, NIT Jalandhar as performance guarantee amount and hereby irrevocably, unconditionally and absolutely undertake against any loss or damage caused or suffered by NIT Jalandhar by reason of any failure of the supplier to perform or omission or negligence to perform any part of its obligations to the satisfaction of NIT Jalandhar in terms of the order.

We, the supplier, do hereby authorize Registrar, NIT Jalandhar to forfeit this Performance Guarantee amount / undertake to pay the amount due and payable under this guarantee without any demur merely on a demand from the NIT Jalandhar stating that the amount claimed is due by way of loss or damage caused to or would be caused to or suffered by the NIT Jalandhar by reason of any breach by us of any of the terms and conditions contained in the said order or by reason of our failure or omission or negligence to perform the said order or any part thereof. We, the Supplier, undertake to pay to NIT Jalandhar any amount so demanded by NIT Jalandhar, notwithstanding:

- a) Any dispute or difference between NIT Jalandhar and supplier or any other person or between the supplier or any person or any suit or proceeding pending before any court or tribunal or arbitrator relating thereto; or
- b) The invalidity, irregularity or unenforceability of the order; or
- c) Any other circumstances which might otherwise constitute discharge of this guarantee, including any act of omission or commission on the part of NIT Jalandhar to enforce the obligations by the supplier or any other person for any reason whatsoever.

We, the Supplier, further agree that the performance Bond/ Guarantee herein contained shall be continued one and remain in full force and effect during the period that would be taken for the performance of the said order and that it shall continue to be enforceable till all the dues of the NIT Jalandhar under or by virtue of the said order have been fully paid and its claims satisfied or discharged or till the office of the Registrar, NIT Jalandhar certifies that terms and conditions of the said order have been fully and promptly carried out by us and accordingly discharges this Performance Bond/ Guarantee.

We, the Supplier, further agree with NIT Jalandhar, that NIT Jalandhar shall have the fullest liberty without our consent and without affecting in any manner our obligations hereunder to vary any of the terms and conditions of the said order or to extend time of performance by the said supplier from time to time or to postpone for any time or from time to time and of the powers exercisable by the NIT Jalandhar against the said supplier and forbear or enforce any of the terms and conditions relating to the order and we shall not be relieved from our liability by reason of any such variation or extension being granted to us or for any forbearance, act or omission on the part of NIT Jalandhar or any indulgence by NIT Jalandhar to us or by any such matter or thing whatsoever which under the law relating to sureties would, but for this provision, have effect of so relieving us.

This Performance Bond/Guarantee will not be discharged due to the change in the constitution of the supplier. We, the Supplier, undertake not to revoke this Performance Bond / Guarantee except with the prior consent of NIT Jalandhar in writing.

The disputes relating to this Bank Performance Bond / Guarantee shall be resolved as per the terms and conditions of the order.



FORMAT FOR PERFORMANCE BOND (BANK GUARANTEE)

In consideration for the Registrar, National Institute of Technology Jalandhar, (*hereinafter called NIT Jalandhar*) having agreed to release the payment of net value as per terms and conditions of a concluded Order No. _____ dated _____ (*hereinafter called 'the order'*) for supply of _____ (*hereinafter called 'the Equipment /Item'*) to Messrs _____ (*hereinafter called 'the supplier'*) on submission of a Bank Guarantee to the satisfaction of NIT Jalandhar for the due performance of the said order.

We, _____ (*hereinafter called 'the Bank'*) at the request of the supplier do, as a primary obliger and not merely as surety, hereby irrevocably, unconditionally and absolutely undertake against any loss or damage caused or suffered by NIT Jalandhar by reason of any failure of the supplier to perform or omission or negligence to perform any part of its obligations to the satisfaction of NIT Jalandhar in terms of the order.

We, the Bank do hereby undertake to pay the amount due and payable under this guarantee without any demur merely on a demand from NIT Jalandhar stating that the amount claimed is due by way of loss or damage caused to or would be caused to or suffered by NIT Jalandhar by reason of any breach by the said supplier of any of the terms and conditions contained in a said order or any part thereof. Any such demand made on the Bank shall be conclusive as regards the amount due and payable by the bank under this guarantee, which shall not be considered as satisfied by any intermediate payment or satisfaction of any part of or obligation hereunder. However, our liability under this guarantee shall be restricted to an amount not exceeding _____.

We, the Bank, undertake to pay to NIT Jalandhar any amount so demanded by NIT Jalandhar, notwithstanding a). Any dispute and difference between NIT Jalandhar and supplier or any other person or between the supplier or any person or any suit or proceeding pending before any court or tribunal or arbitrator relating thereto or

- a). The invalidity, irregularity or unenforceability of the order or
- b). Any other circumstances which might otherwise constitute discharge of this guarantee, including any act of omission or commission on the part of NIT Jalandhar to enforce the obligations by the supplier or any other person for any reason whatsoever.

We, the Bank, further agree that the guarantee herein contained shall continue and remain in full force and effect during the period that would be taken for the performance of the said order and that it shall continue to be enforceable till all the dues of NIT Jalandhar under or by virtue of the said order have been fully paid and its claims satisfied or discharged or till the office of the Registrar, NIT Jalandhar confirms that the terms and conditions of the said order have been fully and promptly carried out by the said supplier and accordingly discharge this guarantee.

We, the Bank, hereby agree and undertake that any claim which the bank may have against the supplier shall be subject to and subordinate to the prior payment and performance in full of all the obligations of the bank hereunder and the bank will not, without prior written consent of NIT Jalandhar, exercise any legal rights or remedies of any kind in respect of any such payment or performance so long as the obligations of the bank hereunder remain owing and outstanding, regardless of the insolvency, liquidation or bankruptcy of the supplier or otherwise. We, the Bank, will not counter claim or set off against its liabilities to NIT Jalandhar hereunder any sum outstanding to the credit of NIT Jalandhar with it.

We, the Bank, further agree with NIT Jalandhar, that NIT Jalandhar shall have the fullest liberty without our consent and without affecting in any manner our obligations hereunder to vary any of the terms and conditions of the said order or to extend time of performance by the said supplier from time to time or to postpone for any time or from time to time and of the powers exercisable by the NIT Jalandhar against the said supplier and forbear or enforce any of the terms and conditions relating to the order and we shall not be relieved from our liability by reason of any such variation or extension being granted to the said supplier or for any forbearance, act or omission on the part of NIT Jalandhar or any indulgence by NIT Jalandhar to the said supplier or by any such matter or thing whatsoever which under the law relating to sureties would, but for this provision, have effect of so relieving us.

This guarantee will not be discharged due to the change in constitution of the Bank or the supplier.

We, the Bank, lastly undertake not to revoke this Guarantee during its currency except with the prior consent of NIT Jalandhar in writing.

The disputes relating to this Bank Guarantee shall be resolved as per the terms and conditions of the order.



FORMAT FOR MANUFACTURER'S AUTHORIZATION FORM

To,
The Registrar
Dr B. R Ambedkar National Institute of Technology
Jalandhar

Sub. : e-Tender for "_____".

Dear Sir,

We, _____, who are established and reputed manufacturers of _____, having factory/office at _____, hereby authorize M/s _____ [name & address of agents/distributors] to bid, negotiate and conclude the order with you for the above goods manufactured by us.

We shall remain responsible for the tender/Agreement negotiated by M/s _____, jointly and severally. No company or firm or individual other than M/s _____ are authorized to bid, negotiate and conclude the order in regard to this business against this specific tender as for all business in the entire territory of India.

We hereby extend our full guarantee and warranty as per the terms and conditions of tender for the goods offered for supply against this invitation for bid by the above supplier.

****specify in detail manufacturer's responsibilities the services to be rendered by***

M/s _____ are as under:

- i) _____
- ii) _____

[Specify the services to be rendered by the agent/distributor] In case duties of the agent/distributor are changed or agent/ distributor is changed it shall be obligatory on us to automatically transfer all the duties and obligations to the new Indian Agent failing which we will ipso-facto become liable for all acts of commission or omission on the part of new Indian Agent/ distributor.

Yours faithfully,

[Name & Signature] For and on behalf of M/s. _____ [Name of manufacturer]



Annexure "E"

DECLARATION REGARDING BLACKLISTING / DEBARRING FOR TAKING PART IN TENDER

Self Attested

I / We _____ (Tenderer) hereby declare that the firm / agency namely M/s. _____ has not been blacklisted or debarred in the past by Union / State Government or organization from taking part in Government tenders in India.

Or

I / We _____ (Tenderer) hereby declare that the Firm / agency namely M/s. _____ was blacklisted or debarred by Union / State Government or any Organization from taking part in Government tenders for a period of

_____ years w.e.f. _____ to _____. The period is over on _____ and now the firm/company is entitled to take part in Government tenders.

In case the above information found false I / we are fully aware that the tender/ contract will be rejected / cancelled by Director, NIT Jalandhar, and EMD / SD shall be forfeited.

DEPONENT

Attested:

(Stamp of Company with authorized sign)

Name _____

Address _____



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY

G T Road By Pass, Jalandhar-144011, Punjab (India)

EPABX-0181-2690301 & 453 website: www.niti.ac.in email: registrar@niti.ac.in

Annexure: "F"

CERTIFICATE OF WARRANTY

- i). We certify that the warranty shall be for a period of _____ years for _____ and starting from the date of satisfactory installation, commissioning and handing over of the Equipment /Item and of the works conducted therewith covered under the supply order in working order. During the **warranty period, I/we shall provide free "after sale service" and the replacement of any part(s) of the Equipment /Item** or rectification of defects of work of the Equipment /Item will be free of cost. The replacement of the parts shall be arranged by us, at our own cost and responsibility. We undertake that the above warranty shall begin only from the date of satisfactory and faultless functioning of the Equipment /Item for 60 days at NIT Jalandhar premises. The benefit of change in dates of the warranty period shall be in the interest of the use/your organization.
 - ii). During the warranty period, we shall provide at least _____ **preventive maintenance visits**.
 - iii). Uptime Guarantee: During the warranty period, we will be responsible to maintain the Equipment /Item in good working conditions for a period 350 days (i.e. 95% uptime) in a block of 365 days.
 - a). All complaints will be attended by us within 2 working days of receipt of the complaint in our office.
 - b). In case there is delay of more than 2 days in attending to a complaint from our side then you can count the number of days in excess of the permissible response time in the downtime. The above said response time of 2 days for attending to a complaint by us will not be counted in the downtime.
 - c). **Penalty:** We shall pay a penalty equivalent to **0.5 % of the FOB/CIF** value of the Equipment /Item for every week or part thereof delay in rectifying the defect.
- Note: The right to accept the reason (s) for delay and consider reduction or wave off the penalty for the same shall be at the sole discretion of Director, NIT Jalandhar**
- iv). We certify that the Equipment /Item being/ quoted is the latest model and that spares for the Equipment /Item will be available for a period of at least _____ years and we also guarantee that we will keep the organization informed of any update of the Equipment /Item over a period of _____ years.
 - v). We guarantee that in case we fail to carry out the maintenance within the stipulated period, NIT Jalandhar reserves the right to get the maintenance work carried out at our risk, cost and responsibility after informing us. All the expenses including excess payment for repairs/maintenance shall be adjusted against the Performance Bank Guarantee. In case the expenses exceed the amount of Performance Bank Guarantee, the same shall be recoverable from us with/without interest in accordance with the circumstances.
 - vi). We shall try to repair the Equipment /Item at NIT Jalandhar premises itself. However, the Equipment /Item will be taken to our site on our own expenses in case it is not possible to repair the same at NIT Jalandhar. We shall take the entire responsibility for the safe custody and transportation of the Equipment /Item taken out for repairs till the Equipment /Item is rehabilitated to the NIT Jalandhar after repair Any loss of Equipment /Item or its accessories under its charge on account of theft, fire or any other reasons shall be at our sole risk and responsibility which will be compensated to NIT Jalandhar for such losses at the FOB/CIF value for the damaged/lost Equipment /Item part, including accessories.
 - vii. We undertake to perform calibration after every major repair/breakdown/taking the Equipment /Item for repair out of NIT Jalandhar premises.
 - viii. In case of extended warranty, we undertake to carry out annual calibration of the Equipment /Item.
 - ix. We guarantee that we will supply spare parts if and when required on agreed basis for an agreed price. The agreed basis could be an agreed discount on the published catalogue price.
 - x. We guarantee to the effect that before going out of production of spare parts, we will give adequate advance notice to you so that you may undertake to procure the balance of the life time requirements of spare parts.
 - xi. We guarantee the entire unit against defects of manufacture, workmanship and poor quality of components.

Signature & Seal of the Manufacturer/Tenderer



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY

G T Road By Pass, Jalandhar-144011, Punjab (India)
EPABX-0181-2690301 & 453 website: www.niti.ac.in email: registrar@niti.ac.in

Annexure: "F"

SELF DECLARATION [For Local Content of Products, Services or Works]

To,

The Director
Dr.B.R.Ambedkar NIT Jalandhar

1. With reference to Order No. P- 45021/2/2017-PP(BE-II) dated 16-09-2020 of DPIIT, Ministry of Commerce and Industry, Govt. of India, we fall under the following category of supplier (please tick the correct category) for the items for which this tender has been floated and being bided.

 Class I local supplier – has local content equal to more than 50%. Local contents added at _____(name of location).

 Class II local supplier – has local content more than 20% but less than 50%. Local contents added at _____(name of location).

 Non-local supplier – has local content less than or equal to 20%. Local contents added at _____ (name of location).
2. We are solely responsible for the abovementioned declaration in respect of category of supplier. False declarations will be in breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which we may can be debarred for up to 2 years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under law.

Signature & seal of the company

Name and address of the organization

.....
.....
.....

Note: In cases of procurement for a value in excess of Rs. 10 crores, the 'Class-I local supplier' / 'Class-II local supplier' shall be required to provide a certificate from the statutory auditor or cost auditor of the company (in the case of companies) or from a practicing cost accountant or practicing chartered accountant (in respect of suppliers other than companies) giving the percentage of local content.