



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY
G T Road By Pass, Jalandhar-144011, Punjab (India)
Tel: 0181-2690301-2690453 PBX Fax: 2690320 website www.nitj.ac.in

Ref. No: NITJ/Pur/PH/144/17/

Date:

NOTICE INVITING QUOTATION

Sealed quotations are invited on behalf of Director, NIT, Jalandhar for the Supply of Different Equipment required for Solid State Lab of Physics Department.

Please send your offers super scribing

- (i) “Quotation for Four Hall Effect Set – 01 No
Trace Hysteresis – 01 No
Gauss Meter - 01 No
(ii) Reference of this letter/ NIQ No.
(iii) Opening date of quotation on the cover of the envelope.

Last date of receipt of quotation	09.07.2018 at 3:00 PM
Opening of quotation	09.07.2018 at 3:30 PM

Quotations should reach at following address before the last date of receipt and time.

*Kind Attention: Purchase Section,
Director*

*Dr. B.R. Ambedkar National Institute of Technology, Jalandhar
GT Road, By Pass, Jalandhar-144001*

The quotations will be opened by the committee duly constituted for the purpose on due date and time in the presence of the quotees, whom so ever may wish to be present.

1. All the quotations/enquiries should be addressed to the Director, Dr B R Ambedkar National Institute of Technology, Jalandhar.
2. In case of non –compliance of instructions and all the terms & conditions, the offer will not be considered.
3. The institute being an Educational Institution, a special rebate/discount may be allowed.
4. Quotations without GST number will be invalid.

(Download Detailed Terms & Conditions, Specification of above equipment as per Annexure-I &II from institute’s website www.nitj.ac.in.)



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY
G T Road By Pass, Jalandhar-144011, Punjab (India)
Tel: 0181-2690301-2690453 PBX Fax: 2690320 website www.nitj.ac.in

Ref. No: NITJ/Pur/PH/144/17/

Date:

IMPORTANT NOTE

The offers must be submitted in a sealed envelope duly super scribing the Quotation for Sealed quotations are invited on behalf of Director, NIT, Jalandhar for the Supply of Different Equipment required for Solid State Lab of Physics Department with due date and time on the cover of envelope.

1. The Quotations received late, will not be considered.
2. All corrigenda, addenda, amendments and clarifications to Notice Inviting Quotation will be hosted in the website www.nitj.ac.in. Quotee shall keep themselves updated with all such developments.
3. In case the last date of receipt/ opening of Quotation falls on holiday, the bids shall be opened on the following working day at same time.
4. The Quotation should be neatly typed. Any deviation in the offer shall lead to rejection.
5. Quotation(s) received beyond last date of offer submission will be rejected. No offer will be entertained by Email/FAX.
6. Supplier must mention Full address with Telephone No., Fax No., GST No. & E mail on their envelop also.



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY
G T Road By Pass, Jalandhar-144011, Punjab (India)
Tel: 0181-2690301-2690453 PBX Fax: 2690320 website www.nitj.ac.in

Ref. No: NITJ/Pur/PH/144/17/

Date:

Annexure-I

**FOLLOWING TERMS AND CONDITIONS MAY BE CAREFULLY STUDIED BEFORE
SUBMITTING QUOTATIONS**

1. Name, complete address of the bidder/Quotee with telephone No., Fax No. ,
GST No. & Email.

2. Location of Head office with complete address, with telephone No. Fax No. and
Email.

3. The quotation must be quoted duly typed on the letter heads of the supplier /
manufacturer duly mentioned full address ,GST No./ PAN No./ and must be in a
sealed cover which should be super scribed Quotation for

_____ Quotation No. NITJ/Pur/ dt:_____ ”.

4. Telefax /E-mail and other incomplete offer will not be considered.

Read and Accepted

**Signature of the Quotee
With Stamp**



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY

G T Road By Pass, Jalandhar-144011, Punjab (India)

Tel: 0181-2690301-2690453 PBX Fax: 2690320 website www.nitj.ac.in

Ref. No: NITJ/Pur/PH/144/17/

Date:

5. All quotations will be deemed to be FOR NIT, Jalandhar unless otherwise specified and should include packing and forwarding charges.
6. If quoting otherwise than FOR NIT, Jalandhar, insurance, charges from place of loading to destination shall be **borne** by the supplier.
7. All quotations will be valid for at least 90 days excluding the date of opening
8. **Illustrated literature and catalogue must be furnished along with quotation.**
9. **The supplier must intimate whether he is manufacturer, dealer or supplier. In case of supplier or dealer the authorization from the manufacturer is required and copy of the same must be enclosed with the bid/quotation.**
10. Rate should be quoted in the Indian rupees only. Delivery period must be mentioned.

Read and Accepted

**Signature of the Quotee
With Stamp**



Ref. No: NITJ/Pur/PH/144/17/

Date:

11. **Quotation received after the due date will not be considered unless the date is extended.**
12. The successful supplier has to submit an amount of @2% of the total value of the offer in shape of bank guarantee/EMD upto the period of warranty period of the equipment before releasing of 10% payment **(for equipment which costs more than Rs. 1,00,000/- individually).**
13. In case the last date of receipt of quotation falls on holiday, the quotation shall be opened on the following working day.
14. **The make/model and country of origin must be specified.**
15. Freight paid sample must reach this office before or at the latest by the due time and date of the opening of the quotations, if required.
16. It may be stated clearly whether the items quoted are available in stock. If not the minimum period of supply of goods be indicated.
17. **Percentage of taxes to be charged must be clearly stated.**
18. All containers will be deemed to be non-returnable, unless specifically stated in the quotations.
19. The approved samples will be sealed duly signed by the representative of the quote and kept in the office at the time of release of the order (where applicable)
20. Inspection of goods will be carried out by the consignee or the Inspection Committee at the destination and rejected goods will have to be removed within 10 days from the date of dispatch of advice from the indenting officer, failing which the quote shall be responsible for the loss.

Read and Accepted

**Signature of the Quotee
With Stamp**



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY
G T Road By Pass, Jalandhar-144011, Punjab (India)
Tel: 0181-2690301-2690453 PBX Fax: 2690320 website www.nitj.ac.in

Ref. No: NITJ/Pur/PH/144/17/

Date:

21. The rejected goods must be replaced within 15 days of the dispatch of advice by the indenting officer or a registered notice, failing which indenting officer will be entitled to make arrangements of its disposal without any further reference.
22. If there is any dispute/doubt, the decision of the Director, Dr. B. R. Ambedkar National Institute of Technology, Jalandhar shall be final.
23. All disputes are subject to Jalandhar Jurisdiction.
24. Certificate to the effect that the Price quoted by you will be lowest and is not more than the Price charged from other Educational Institutions/ R & D departments in India.
25. Copies of supply order at least of three users where you have already supplied above said Equipment/items/software be supplied.
26. Details of training and the cost if any in the institute be intimated.
27. The supplier who quote the equipment, must provide terms & conditions, upgrade policy and installation etc. and maximum discount can be allowed to the Institute being an Educational Institution be supplied.
28. **90% payment will be released against delivery and inspection of the material and balance 10% payment shall be released within 20 days after having the inspection and installation report of the equipment/material. Execution of the complete order of goods as per specification of the supply order and their installation and working.**

Read and Accepted

**Signature of the Quotee
With Stamp**



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY
G T Road By Pass, Jalandhar-144011, Punjab (India)

Tel: 0181-2690301-2690453 PBX Fax: 2690320 website www.nitj.ac.in

Ref. No: NITJ/Pur/PH/144/17/

Date:

29. The quantity of material may vary as per requirement.
- 30 **Penalty:** The Director, Jalandhar reserves the right to impose 0.5% (Half) per cent penalty per week on account of delay in supply, if delivery received after expiry of the original delivery period. The total penalty should not exceed 10% of the value of the delayed goods.

Read and Accepted

**Signature of the Quotee
With Stamp**



Ref. No: NITJ/Pur/PH/144/17/

Date:

Annexure-II

Technical Specification

Hall Effect Setup

Hall Effect Setup that consists of :

(1) **ELECTROMAGNET 7.5KG.** Its Pole pieces distance is 10mm and diameter of pole pieces is 75mm.

(2) **Digital DC Power Supply 32 Volts/5Amps.**

(3) **Hall Probe**

Ge Crystal N & P –Type with Four Spring type pressure contacts is mounted on sunmica decorated Bakelite strip .Four leads are provided for connection with measuring devices.

Hall Voltage: 0.1-1 Volts / 100Ma per KG.

Thickness of Ge Crystal: 0.4 to 0.5mm

Resistivity : approx.. 10ohmcm

(4) Instructional Lab Manual



Ref. No: NITJ/Pur/PH/144/17/

Date:

Technical Specification

B-H Trace Hysteresis curve Apparatus

B-H Trace Hysteresis curve Apparatus that consists of

- (1) 12 Volts variable AC Power Supply inbuilt
- (2) Band Switch (BS1) having resistance of ($R \times 10k\Omega$) on the Front Panel
- (3) Band Switch (BS2) having capacitances of ($0.1\mu f$, $0.2\mu f$, $0.3\mu f$, $0.4\mu f$, $0.5\mu f$) on the front panel
- (4) THREE output banana sockets (4mm) marked as X-Probe, GND and Y-Probe are given on the front panel for the CRO connection.
- (5) A Specimen transformer (0-12V) is fixed inside trainer kit.
- (6) Two connecting probes (Banana Plug(4mm) one end BNC Socket on other end) to connect the output Terminals of B-H curve circuit with CRO.
- (7) Instructional Lab Manual.



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY
G T Road By Pass, Jalandhar-144011, Punjab (India)

Tel: 0181-2690301-2690453 PBX Fax: 2690320 website www.nitj.ac.in

Ref. No: NITJ/Pur/PH/144/17/

Date:

Technical Specification

Digital gauss meter :

(1)Range: 0-2K Gauss AND 0-20K Gauss.

Digital gauss meter that consists of :

(2)Resolution : 1gauss at 0-2K gauss range.

(3) Display: 3.5 Digits seven Segments LRD display

(4) Power: 220, 50Hz

Gauss Probe

In-As Crystal (Rectangular Type) is mounted on phenolic strip with four soldered contacts for connection with measuring devices.

Hall VOLTAGES: 8-10 mV / 100Ma per Kg.