CORRIGENDUM

Subject: - Corrigendum Ref. No NITJ/PUR/CE/e-Tender No.09/2019

As technical specification of the required equipment (Jaw Crusher) has undergone slight technical modification, the revised technical specification may please be seen as per the annexure 1. Bidders are accordingly requested to apply as per revised specifications only.

Registrar



Dr B R AMBEDKAR NATIONAL INSTITUTE OF TECHNOLOGY JALANDHAR

Annexure-T

Revised Technical Specifications of <u>Jaw Crusher</u>

S.No.	Specifications
1.	Vendor has to quote for the Jaw crusher
2.	The Jaw crusher should be able to crush hard materials
3.	Gap width setting feature should be there in the Jaw crusher
4.	Final fineness of the material should be up to < 2mm
5.	Jaw crusher should be installed with zero point adjustment
6.	The minimum collector capacity for the Jaw crusher should be 4 litre
7.	The feature of over load protection should be installed in the Jaw crusher
8.	The Jaw crusher should be able to crush wide range of materials so that contamination free grinding should takes place
9.	Separate dust connector should be installed in the crusher for dust extraction
10.	The Jaw crusher should be suitable for integration in automatic installations
11.	Heavy duty manganese steel or stainless steel should be used in the grinding tools
12.	The instrument must meet the CE /international safety standards
13.	The feed size for the Jaw crusher should be more than 80mm
14.	The throughput/output of the Jaw crusher should be at least 200kg/hr.

Dr. Kanish Kapoor Assistant Professor Dr. S P Singh Professor